



Microbix Biosystems Inc.

265 Watline Avenue
Mississauga, Ontario, Canada L4Z 1P3
Tel 905-361-8910 Fax 905-361-8911

Product Name:

Adenovirus grade 2 Antigen

Catalogue Number: EL-15-02

Storage: Store this antigen preparation frozen at - 70 °C to - 100 °C. Repeated freezing and thawing should be avoided.

Hazards: We are aware of no specific hazards associated with this product. The reagent has been inactivated and should contain no infectious material. Generally accepted good laboratory practices appropriate to biological reagents should be employed when handling this product.

Strain: Adenoid 6

Cultured In: MRC-5

Buffer: MEM

Agent Description: Adenoviridae family. A non enveloped, icosahedral virion, 75-80 nm diameter, doubl-stranded, linear DNA genome. Varies in clinical manifestation and severity. Symptoms include fever, rhinitis, pharyngitis, cough and conjunctivitis.

Preparation: Optimally infected cells are disrupted by sonication and then subjected to low speed centrifugation. The clarified cell lysate is pooled with supernatant from the infected culture and concentrated using crossflow ultrafiltration with a 30,000 molecular weight cutoff.

Inactivation: Adenovirus antigen is inactivated using gamma radiation. This procedure is effective primarily by damaging viral genetic material.

Description: The antigen preparation contains a high concentration of virus and viral components as well as some cellular material in MEM.

Recommendations for Use: This antigen preparation should be sonicated immediately prior to use to ensure that the preparation is uniform. This preparation may be used as is in a variety of immunoassay formats or may be further purified to meet the needs of a particular assay format. Grade 2 antigen is widely used for both IgG and IgM detection in assays which include EIA with polystyrene and latex solid phases.

Assistance: If you have any questions regarding the production, testing or use of this antigen, please send them by email to customer.service@microbix.com or fax 905-361-8911, with any relevant data, to Microbix Technical Services. Your complete satisfaction with the performance of this product is important to us.

Quality Control Information

Product Name: Adenovirus grade 2 Antigen

Lot Number: 15014B1

Microbix performs quality control tests to ensure each batch meets in-house specifications. Test results are provided with each lot of antigen shipped. Antigen users require this information for a number of reasons:

- to maintain a record for good manufacturing purposes,
- to correlate user results with Microbix results and
- for use as a starting point for those just starting with either a new antigen or developing a new assay.

It is important that each user perform titrations of antigen using their own assay as each assay format and serum release panel makes different performance demands on the antigen. Often, use of an antigen may be optimized by making adjustments to concentrations of other assay reagents such as conjugate. Once this is complete the result is cost effective use of the antigen and optimal assay performance.

Test:

Titre: This antigen preparation is titrated using a microtitre plate based ELISA. (This procedure may be found in Microbix Technical Bulletin number 93-1.) Antigens are tested for reactivity with IgG. The dilution of antigen which generates a signal of 1.0 O.D. unit in the immunoassay is compared to that of the standard approved antigen. The result of this comparison is expressed as a percentage of the reference.

IgG Result: 97.5 %

Protein Concentration: Protein is determined using the Biorad dye binding assay in the microassay format. The standard curve is generated with a known concentration of IgG.

Result: 3.16 mg/ml

Inactivation Assay: The effectiveness of inactivation is tested by inoculating a MRC-5 monolayer with antigen. The culture is manipulated using the original optimal culture conditions used to manufacture the antigen. The culture is monitored for cytopathic effect for 5 days. If no sign of infection is observed the culture is passaged into a fresh monolayer. The second passage is monitored for a further 5 days. If no cytopathic effect is observed in either passage the antigen is considered inactivated.

Result: No growth detected



Quality Assurance Signature:

23-February, 2011